सी एस आई आर -राष्ट्रीय पर्यावरण अभियांत्रिकी अनुसंधान संस्थान

CSIR- National Environmental Engineering Research Institute नेहरु मार्ग, नागपुर ४४००२०, भारत



ईमेल Email:- spo@neeri.res.in, st_pur@neeri.res.in

वैबसाइटWebsite: www.neeri.res.in

CORRIGENDUM

PUR-85/EP/SEAF/2022-23

Corrigendum is issued for following correction in the Tender Document:

- "1) Technical specifications of Microwave Digestion System is amended and revised specification is as per Annexure-A.
- 2) Bid submission end date may be read as 27.10.2022 instead of 18.10.2022
- 3) Date of Opening of the bid may be read as **28.10.2022** instead of 19.10.2022."

Qualification criteria, Terms & Conditions will remain same.

भंडार एवं क्रय अधिकारी Stores & Purchase Officer

Date: 07.10.2022

Revised Technical Specification for Microwave Digestion System.

Subject: Revised Technical Specification for the Purchase of Microwave Digestion System

Supply, installation, demonstration and testing of Front loaded table top Microwave Digestion System with all required accessories and operating software

MICROWAVE DIGESTION SYSTEM (FRONT LOADED FOR SAFETY ISSUES with

<u>SS solid door</u>) with 24 Nos. vessel rotor for quick dissolution of various environmental samples without loss of heavy metals from sample solution. Rotor should be upgradable.

The system should allow pressurized reaction with user selection of continuous or pulsed microwave reaction.

The system should have metal body and should have solid steel door for safety.

MAGNETRON:

The microwave system should have dual Magnetron system with good diffuser for homogenous microwave power distribution in the cavity. Microwave frequency should be 2455 MHz and installed power should be minimum 1800 W (two magnetrons minimum 900W each). The delivered / output power should be controllable via microprocessor +/- 1W increments. There should be provision for protection from reflected microwave energy. The microwave delivery should be selective pulsed or un-pulsed depending upon the user.

CAVITY:

The cavity should be made of stainless-steel housing with PTFE plasma coating for corrosion resistance. Also, all hardware should have protective coating for the resistance from acid. The microwave cavity should be designed for volume of around 50 litre or more. There should be provision of door lock so that it cannot be opened automatically if the temperature inside the vessel is not safe.

SAFETY:

- Interlocks: The system should have good interlocking system for safety and cavity door should have 3 or more safety interlocks.
- The system door should have interlock so that unless inside vessel temperature is below 50 °C door should not open automatically.
- Vessels: Vessels offered should have venting technology in case of excess pressure development.
- Safety Certification: The system should be certified by necessary safety authority for safe operations by any international agencies.

EXHAUST:

The system should have inbuilt exhaust system to cool the vessels and to drive away if any fumes in the cavity and it should cover under three year warranty.

VESSELS:

The system should be high throughput which can hold at least 24 Nos. The material of construction should be high purity TFM fluoropolymer. Vessel volume should be about 80 ml or more and it should be designed for maximum temperature capacity upto 250°C and maximum pressure 60 Bar.

SENSORS:

The system should have IR temp sensor for each vessel temp, controlling and monitoring and contactless pressure sensor or accurate pressure controlling mechanism for all vessels should be offered with the system

CONTROLLER:

System should be controlled by touchscreen display.

POWER:

Power requirement: 230V, 50 Hz

WARRANTY:

One Year Standard Warranty + 2 year additional warranty